



659 Andover Street
Lawrence, MA 01843-1032, USA
Tel: 1.978.687.9200 Fax: 1.978.687.9216

Certificate of Analysis

Product Name: Viral Transport Medium

Lot Number: VTM-0011
VTM-0012
VTM-0013
VTM-0014
VTM-0015
VTM-0016
VTM-0017

Expiration Date: 06 May 2021

Order Code: CDC-VTM

Ingredients: Viral Transport Medium contains;
2% Fetal Bovine Serum
Hanks Balanced Salt Solution
100 µg/mL Gentamicin
0.5 µg/mL Amphotericin

Analysis:

Quality Control Test	Specification	Result
Sterility (No visual turbidity)	Clear, colorless liquid	Pass
Sterility (Bacterial) – Blood Agar Plate	No growth	Pass
Sterility (Yeast/Mold) – Peel Plate® Y/M	No growth	Pass
SARS-CoV-2 (Coronavirus) RNA-Free	Negative	Pass
Viral Transport Medium Volume	2.0 ± 0.5 mL	Pass
Vial Leak Check	No leakage	Pass

This product was quality tested and met Charm Sciences' quality control criteria for release for sale including performance and appearance specifications.

Name: David R. Legg
Title: Vice President, Quality Assurance
Date: 15 May 2020