



659 Andover Street  
Lawrence, MA 01843-1032, USA  
Tel: 1.978.687.9200 Fax: 1.978.687.9216

## Certificate of Analysis

Product Name: Viral Transport Medium

Lot Number: VTM-0105  
VTM-0106

Expiration Date: 15 July 2021

Order Code: CDC-VTM, CDC-VTM-100K

Ingredients: Viral Transport Medium contains;  
2% Fetal Bovine Serum  
Hanks Balanced Salt Solution  
100 µg/mL Gentamicin  
0.5 µg/mL Amphotericin

### Analysis:

Quality Control Test	Specification	Result
Sterility (No visual turbidity)	Clear, colorless liquid	Pass
Sterility (Bacterial) – Blood Agar Plate	No growth	Pass
Sterility (Yeast/Mold) – Peel Plate® Y/M	No growth	Pass
SARS-CoV-2 (Coronavirus) RNA-Free	Negative	Pass
Viral Transport Medium Volume	3.2 ± 0.3 mL	Pass
Viral Transport Medium pH	7.4 ± 0.3	Pass
Vial Leak Check	No leakage	Pass

This product was quality tested and met Charm Sciences' quality control criteria for release for sale including performance and appearance specifications.

Name: David R. Legg  
Title: Vice President, Quality Assurance  
Date: 25 August 2020